$\frac{Format\,\text{-B2}}{\text{REPORT ON ORDERS AWARDED ON NOMINATION BASIS}}\;\;\text{FOR THE MONTH, JULY' 2015}$

Sphere: Duliajan Department : Materials

SI	L. NO	Department	Name of the work and location	Tender cost Rs. (Lakhs)	Agreement No. and date of the award	Name of the Agency to whom contract awarded	Date/letter No. to Board for scrutiny and vetting	Remark of the board	Remarks of the department (conditions which necessitated nomination)
	1	2	3	4	5	6	7	8	9
	1	Field Engineering	Supply, Servicing & Calibration of pH meter.	1.85	7510562/N5 dtd. 29.07.2015	M/s POWERGRIPP UNIVERSAL, Sivasagar	None	None	To manintain the quality of treated water there are three on-line meters installed in the Water Treatment Plant for monitoring the Turbidity, pH and Chlorine residue treated in water. The on-line meters are presently malfunctioning and not working properly. It had been subsequently taken up with the Chemical Deptt. to rectify the problem, however it could not be done as these need to be done thoroughly serviced before calibration. In the meantime, the matter was taken up with the supplier of the equipment, M/s Powergripp Universal, Sivasagar for one time servicing, calibration & certification for the same. The supplier has visited the above installation and said that both servicing & calibration are required to put them into normal operation again. They have forwarded their quotation along with the solutions and certification of the equipments. Since the above meters are essential for on-line monitoring of parameters of the treated water, approval of competent authority vide prop. no. FE/W.1.0/323/2015 dtd. 10.03.2015 obtained for servicing calibration and certification of online-meters on nomination basis from M/s Powergripp Universal, Sivasagar.

Certification:-

All works/Purchase/Consultancy and other contracts, in progress, as per prescribed monetary limit and approved under Srl. No. 34 I(A) of DOP have been reported in this QPR.