

OIL INDIA LIMITED
(A Government of India Enterprise)
P.O. Duliajan, Pin – 786602
Dist-Dibrugarh, Assam

AMENDMENT NO. 3 DATED 30.04.2016
TO TENDER NO SDG9753P16/07

This Amendment no. 3 dated 30.04.2016 to tender no. SDG9753P16/07 is issued to amend Technical Specification as under: –

AMENDMENTS IN TECHNICAL SPECIFICATIONS FOR TENDER NUMBER SDG9753P16/06 FOR SUPPLY AND COMMISSIONING OF HPLC SYSTEM

Clause Number	Tender Clauses	Amended Clause
10	Flow rate range from 0.001 to 30 ml/min	0.2 ml or lower to 10ml/min or higher
	Flow cell Design and cell volume : Taper Slit; 9.3ul	Flow cell Design and cell volume : Taper Slit; 8ul or more
	Temperature Control 30-55 °C: 80 °C Thermal protector	Temperature Control 30-50 °C or higher end
	Offset Range: 0-500 mV	Please drop this clause
	Display: Colour liquid crystal display	Display: Liquid crystal display (LCD)/Light emitting diode (LED)
	Column capacity: Upto 12 columns, depending on column length	minimum Six columns, depending on column length
	Eluent pre-conditioner 11ul	Please drop this clause

All other terms and condition of the tender remain unchanged.

DGM-MATERIALS