

OIL INDIA LIMITED
(A Government of India Enterprise)
P.O. Duliajan, PIN: 786602
Dist.: Dibrugarh, Assam

CORRIGENDUM

Amendment No. 1 dated 26.04.2019 to Tender No. SDI0759P19

This Amendment No. 1 dated 26.04.2019 to the Tender No. SDI0759P19 dated 05.03.2019 for 'Supply of Operation Theatre Light' is issued to amend the following:

1. To amend the material description for **item no. 10** of Annexure-IA of the Tender as under instead of the existing:

Clause No	Existing Material Description	Amended Material Description
Material Description of Item 10 (Annexure IA)	<p>Specification for OT Light</p> <p>Should be an internationally reputed brand with CE certification/USFDA</p> <p>* Should be a dual dome with the combination of more than equal 200,000 lux (main dome) + more than equal 200,000 lux (satellite dome intensity).</p> <p>* Each dome should consist of 90 LED's or more.</p> <p>* Each dome should have white LED's with advanced Lens based technology and should be a glare free light.</p> <p>* Domes should be of aerodynamic design to suit laminar air flow operation theatres.</p> <p>* The light beam should be homogeneous with the depth of cavity penetration of at least 120 cm for each dome.</p> <p>* Should have high performance LED's with at least 50,000 hours service life.</p> <p>* Should have a variable colour temperature of 3500K-5000K.</p>	<p>Amended Specification for OT Light:</p> <p>Should be an internationally reputed brand with CE certification / USFDA</p> <p>* Should be a dual dome with the combination of more than equal 1,60,000 lux (main dome) + more than equal 1,60,000 lux (satellite dome intensity).</p> <p>* Each dome should consist of 90 LED's or more.</p> <p>* Each dome should have white LED's with advanced Lens based technology and should be a glare free light.</p> <p>* Domes should be of aerodynamic design to suit laminar air flow operation theatres.</p> <p>* The light beam should be homogeneous with the depth of cavity penetration of at least 120cm for each dome.</p> <p>* Should have high performance LED's with at least 50,000 hours of service life.</p> <p>* should have a variable colour temperature of 3500K-5000K.</p>

OIL INDIA LIMITED
(A Government of India Enterprise)
P.O. Duliajan, PIN: 786602
Dist.: Dibrugarh, Assam

<ul style="list-style-type: none">* Field diameter should be adjustable through Sterile handle for surgeon convenient.* Field Diameter should be adjustable between 18 cm to 30cm for main dome and satellite dome.* The colour rendering index should be at least 92 or more.* The power consumption of each light head should be less than 120W per dome.* Both the light heads should have unrestricted rotation of 360 degree on its own axis and should allow free movement even in low rooms with unrestricted positioning options.* Temperature rise in operating area should be very less and produced heat should be dissipated for absorbed through a special integrated elements.* Shadow management system adjusted through control Panel.* Authorisation letter from concerned party (manufacturer /Distributor).	<ul style="list-style-type: none">* Field diameter should be adjustable through Sterile handle for surgeon convenient.* Field Diameter should be adjustable between 18 cm to 30cm for main dome and satellite dome.* The colour rendering index should be at least 92 or more.* The power consumption of each light head should be less than 120W per dome.* Both the light heads should have unrestricted rotation of 360 degree on its own axis and should allow free movement even in low rooms with unrestricted positioning options.* Temperature rise in operating area should be very less and produced heat should be dissipated for absorbed through a special integrated element.* Shadow management system adjusted through control Panel.* Authorisation letter from concerned party (manufacturer /Distributor).
---	--

2. To extend the bid closing date of the Tender No. **SDI0759P19** till **09.05.2019.**

All other terms and conditions as per the original Tender No. SDI0759P19 shall remain same. The bidders are requested to refer all amendments along with the original NIT while quoting.

Sd/-
(H. Upadhaya)
Sr. Purchase Officer(IP)
For GM- Materials